

MH Hospital  
**Laboratory Report**

bioMérieux Customer:  
 System #:

Autoprint

Isolate: 05NBC02-1 (Approved)  
 Card Type: GN Bar Code: 2412528503744707 Testing Instrument: 000015F23819 (11496)  
 Setup Technologist: Laboratory Supervisor(LabSuper)

Bionumber: 4041001001500001

Organism Quantity:

**Selected Organism: Achromobacter xylosoxidans**

<b>Comments:</b>	

<b>Identification Information</b>	<b>Card:</b> GN	<b>Lot Number:</b> 2412528503	<b>Expires:</b> Oct 28, 2024 12:00 PKT
	<b>Status:</b> Final	<b>Analysis Time:</b> 6.03 hours	<b>Completed:</b> Jun 4, 2022 15:53 PKT
<b>Organism Origin</b>	VITEK 2		
<b>Selected Organism</b>	99% Probability <b>Achromobacter xylosoxidans</b> Bionumber: 4041001001500001 Confidence: Excellent identification		
<b>Analysis Organisms and Tests to Separate:</b>			
<b>Analysis Messages:</b>			
<b>Contraindicating Typical Biopattern(s)</b>			

<b>Biochemical Details</b>																	
2	APPA	-	3	ADO	-	4	PyrA	+	5	IARL	-	7	dCEL	-	9	BGAL	-
10	H2S	-	11	BNAG	-	12	AGLTp	+	13	dGLU	+	14	GGT	-	15	OFF	-
17	BGLU	-	18	dMAL	-	19	dMAN	-	20	dMNE	-	21	BXYL	-	22	BAlap	-
23	ProA	+	26	LIP	-	27	PLE	-	29	TyrA	-	31	URE	-	32	dSOR	-
33	SAC	-	34	dTAG	-	35	dTRE	-	36	CIT	+	37	MNT	-	39	SKG	-
40	ILATk	+	41	AGLU	-	42	SUCT	+	43	NAGA	-	44	AGAL	-	45	PHOS	-
46	GlyA	-	47	ODC	-	48	LDC	-	53	IHISa	-	56	CMT	-	57	BGUR	-
58	O129R	-	59	GGAA	-	61	IMLTa	-	62	ELLM	+	64	ILATa	-			

Installed VITEK 2 Systems Version: 9.04.4  
 MIC Interpretation Guideline:  
 AES Parameter Set Name:

Therapeutic Interpretation Guideline:  
 AES Parameter Last Modified: