

MH Hospital
Laboratory Report

bioMérieux Customer:

Printed by: LabSuper

System #:

Isolate: 05NBC01-1 (Approved)

Card Type: GN Bar Code: 2412528503741531 Testing Instrument: 000015F23819 (11496)

Setup Technologist: Laboratory Supervisor(LabSuper)

Bionumber: 0043453303500040

Organism Quantity:

Selected Organism: Pseudomonas aeruginosa

Comments:	

Identification Information	Card: GN	Lot Number: 2412528503	Expires: Oct 28, 2024 12:00 PKT
	Status: Final	Analysis Time: 4.83 hours	Completed: Jun 15, 2022 19:38 PKT
Organism Origin	VITEK 2		
Selected Organism	97% Probability Bionumber: 0043453303500040	Pseudomonas aeruginosa Confidence: Excellent identification	
Analysis Organisms and Tests to Separate:			
Analysis Messages:			
Contraindicating Typical Biopattern(s) Pseudomonas aeruginosa URE(16),			

Biochemical Details																	
2	APPA	-	3	ADO	-	4	PyrA	-	5	IARL	-	7	dCEL	-	9	BGAL	-
10	H2S	-	11	BNAG	-	12	AGLTp	+	13	dGLU	+	14	GGT	+	15	OFF	-
17	BGLU	-	18	dMAL	-	19	dMAN	+	20	dMNE	+	21	BXYL	-	22	BAlap	+
23	ProA	+	26	LIP	+	27	PLE	-	29	TyrA	+	31	URE	+	32	dSOR	-
33	SAC	-	34	dTAG	-	35	dTRE	-	36	CIT	+	37	MNT	+	39	5KG	-
40	ILATk	+	41	AGLU	-	42	SUCT	+	43	NAGA	-	44	AGAL	-	45	PHOS	-
46	GlyA	-	47	ODC	-	48	LDC	-	53	IHISa	-	56	CMT	-	57	BGUR	-
58	O129R	-	59	GGAA	-	61	IMLTa	+	62	ELLM	-	64	ILATa	-			

Installed VITEK 2 Systems Version: 9.04.4

MIC Interpretation Guideline:

AES Parameter Set Name:

Therapeutic Interpretation Guideline:

AES Parameter Last Modified:

Page 1 of 2