

MH Hospital  
**Laboratory Report**

bioMérieux Customer:  
 System #:

Printed by: LabSuper

Isolate: 02NBC03-1 (Approved)

Card Type: GP Bar Code: 2422669403009802 Testing Instrument: 000015F23819 (11496)

Setup Technologist: Laboratory Supervisor(LabSuper)

Bionumber: 030402077763231

Organism Quantity: **Selected Organism: Staphylococcus aureus**

<b>Comments:</b>	

<b>Identification Information</b>	<b>Card:</b> GP	<b>Lot Number:</b> 2422669403	<b>Expires:</b> Mar 18, 2025 12:00 PKT
	<b>Status:</b> Final	<b>Analysis Time:</b> 5.85 hours	<b>Completed:</b> Jun 1, 2022 17:34 PKT
<b>Organism Origin</b>	VITEK 2		
<b>Selected Organism</b>	95% Probability <b>Staphylococcus aureus</b> Bionumber: 030402077763231 Confidence: Very good identification		
<b>Analysis Organisms and Tests to Separate:</b>			
<b>Analysis Messages:</b>			
<b>Contraindicating Typical Biopattern(s)</b>			
Staphylococcus aureus URE(2),			

<b>Biochemical Details</b>																	
2	AMY	-	4	PIPLC	-	5	dXYL	-	8	ADH1	+	9	BGAL	+	11	AGLU	(-)
13	APPA	-	14	CDEX	-	15	AspA	-	16	BGAR	-	17	AMAN	-	19	PHOS	+
20	LeuA	-	23	ProA	-	24	BGURr	-	25	AGAL	-	26	PyrA	+	27	BGUR	-
28	AlaA	-	29	TyrA	-	30	dSOR	-	31	URE	+	32	POLYB	+	37	dGAL	+
38	dRIB	+	39	ILATk	+	42	LAC	+	44	NAG	+	45	dMAL	+	46	BACI	+
47	NOVO	-	50	NC6.5	+	52	dMAN	+	53	dMNE	+	54	MBdG	+	56	PUL	-
57	dRAF	-	58	O129R	+	59	SAL	-	60	SAC	+	62	dTRE	+	63	ADH2s	-
64	OPTO	+															

Installed VITEK 2 Systems Version: 9.04.4

MIC Interpretation Guideline:

AES Parameter Set Name:

Therapeutic Interpretation Guideline:

AES Parameter Last Modified:

Page 1 of 2